

## SmERF Bottles July 9, 2014

	Need	Have		
Bottle #	Description	Description	Machine	Correct?
1	Vehicle emissions grade Nitrogen	Nitrogen Vehicle Emission	Zero N2	Yes
2	40% H2/60% He	FID Mix 40% H2/60% N2 (2 bottles)	FID Fuel	No
3	Oxygen	Oxygen compressed	NOx Analyzer & Gas Divider Ozonator Air	Yes
4	90ppm C3H8 (propane) balance Air	90ppm C3H8 (propane) balance N2	THC Analyzer (300ppm range)	No
5	90ppm CO balance N2	90ppm CO balance N2	COL range Span	Yes
6	4.5% CO balance N2	4.5% CO bal N2	COH range span	Yes
7	18% CO2 balance N2	18% CO2 balance N2	CO2 range span	Yes
8	21% O2 balance N2	4500ppm CO 21% O2 Bal N2	O2 range span	Kind of
9	2700 ppm No bal N2	NOTHING...	NOx converter efficiency test	No

Key
Need to order
Need to return
Can be used in O2 analyzer