


engineering A design




ARTICULAR CARTILAGE SMOOTHER FOR ARTHROSCOPIC SURGERIES


THE SCALPEL JOCKEYS

LEEANN HOLD (BE), ETHAN OVERFELT (BE), KATIE FARRIS (BE), SUHAIM ALTALEB (ME), LISSA NICKELL (BE), QINGYUN WANG (EE), QIAN XU (EE)


> SPONSORED BY: DOUGLAS B. HILLER, M.D. **INSTRUCTOR: DEV SHRESTHA, PH.D., P.E.** MENTOR: COLIN BURKHALTER




BACKGROUND CARTILAGE







Courtesy of Douglas B. Hiller, MD





PROBLEM DEFINITION

SURGICAL INTERVENTIONS

Perform total joint replacements

Arthroscopic debridement

PROBLEM STATEMENT

We were tasked with designing a device to address the underlying issues presented when performing arthroscopic debridement. Current shavers utilize a rotating drum-shaped shaving head that, when pressed against cartilage or bone, can cause ripples, which creates a suboptimal post-surgical result.





VALUE PROPOSITION OBJECTIVE:

Our team is driven to optimize arthroscopic debridement by creating a vibrational arthroscopic smoothing device that creates smoother articular surfaces, which results in a post-surgical joint that provides an improved quality of life for patients.

INNOVATIVE GOAL:

To improve the lives of orthopedic surgeons by creating a more ergonomic and functional device than the current technologies.






PROJECT REQUIREMENTS

- **Comfortable and lightweight**
- Shaft diameter of less than 5 mm
- **I** Biologically compatible materials
- **I** Autoclavable
- **Compatible with suction and fluid flow**
- Creates smooth articular surface






APPROACHES CONSIDERED BALL & SOCKET FIXED HEAD



Challenging to manufacture on small scale

Difficult to control and position






Large fillets on edges



MATERIAL SELECTION

Outer suction shaft – Aluminum

- High strength to weight ratio, non-toxic, inexpensive, and sterilizable
- Common in handles and bodies of surgical instrument
- Should be coated to reduce corrosion
- Inner driveshaft and head 316L "medical grade" stainless steel
 - Less expensive than titanium
 - Strong, yet able to be gritted through sandblasting or lapping
 - Common in surgical instruments
 - Corrosion resistant in aqueous and saline environments






GRIT SELECTION UTILIZING CHICKEN BONES AS TEST SUBJECTS

- **Chicken femoral condyles**
- **Grit selection using sandpaper**
 - 300+ produced visibly smoothest result from 8 grits tested












INITIAL IDEAS FOR PRODUCTION OF VIBRATION












FINAL DESIGN - ASSEMBLY

Assembly includes:

- Driveshaft / Smoothing Surface (Pink)
- Outer Shaft (Pink)
- Fluid Chamber
- Connector Jacket (Purple)
- Motor & Clips (Red)
- Trigger / Springs / Push Button (Green)
- Circuit Board (Yellow)
- Battery Compartment (Blue)
- Charger Placement (Orange)
- Nuts & Bolts for fastening
- 3D-Printed Encasing Handle










FINAL DESIGN DETAILS





- **I** Fluid flow facilitated through outer shaft and chamber
- **I** Ergonomic handle design
- **I** Adapter from motor to driveshaft
- **I** Spring-levered trigger control
- Rechargeable battery compartment






ELECTRICAL CIRCUITRY


- Simple code to control vibration
- **I** Ran into problems with button functionality
- Edits resolved problems, now one push of button turns vibration on, another push turns it off

```
int pushbutton;
bool isON;
unsigned long Time;
void setup() {
 pinMode(11,OUTPUT);
 pinMode(12,OUTPUT);
 isON=false;
 pushbutton=0;
 Time = millis();
 attachInterrupt(0,ToggleState,RISING);
void loop(){
  if (isON) {
   digitalWrite(11,HIGH);
   digitalWrite(12,LOW);
   delay(2);
   digitalWrite(11,LOW);
   digitalWrite(12,HIGH);
   delay(2);
void ToggleState() {
   if ((millis() - Time) > 200) {
      isON=!isON;;
      Time = millis();
      Time = millis();
```



ELECTRICAL CIRCUITRY







Vibrational motor

- Simpler than other approaches mentioned
- Microcontroller (Arduino Nano)
- I Period = 4ms, f = 250 Hz, V = 3.7V (from rechargeable battery)
- H-bridge controls waveform of circuit





FINAL PROTOTYPE














SURFACE VALIDATION MORE CHICKEN BONES

I Validation of motor setup and smoothing ability

- Used 40 grit sandpaper to mimic tissue damage
- Used Dremel head to mimic smoothing head of device
- **I** Top to bottom: Healthy tissue, damaged tissue, treated tissue










PROJECT BUDGET





Total Budget: \$3000 Actual Spent: \$1188



FUTURE WORK

- **I** Further testing on suction mechanism
- Use of more advanced microscopy
- Smoothing head design
 - Detachable convex and concave designs
- **Uutsource** Manufacturing
- LONG TERM:
- In vivo trials







engineering 🔺 design







University of Idaho

College of Engineering

For more information:

